

**RECEIVED**  
**CENTRAL FAX CENTER****JUN 15 2009**

Application No.: 10/522,889  
Statement of Substance of the Interview held on May 15, 2009

FREEFLOW MEDICAL LIMITED

Correspondence Address:  
Freeflow Medical Limited  
c/o 123 Deansgate  
Manchester  
M3 2BU  
United Kingdom

E-Mail: rudy.hengelmolen@gmail.com

Application No.: 10/522,889      Confirmation No.: 2152

Applicant and Inventor: Rudy Hengelmolen

Filing Date: 08/18/2005

Technology Centre/Art Unit: 3734

Examiner: Blatt, Eric D

Attorney Docket No.: 1768-135

Mail Date of Office Communication: May 22, 2009

June 14, 2009

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
United States of America

Central fax number: +1 571-273-8300

Total pages transmitted by facsimile on June 15, 2009: 2 pages

**STATEMENT OF SUBSTANCE OF INTERVIEW**

Sir,

In response to the interview held on May 15, 2009, this Statement of Substance of Interview includes a description of discussed features of Applicant's invention as compared to the devices disclosed in Burgermeister and Uflacker.

Examiner asked Applicant to describe the invention as disclosed in the present application. Applicant explained that an important feature of the invention is that the connecting elements are at least twice as long as each of the zig-zag elements since this will improve the flexibility of the device. Examiner acknowledged that upon a cursory review of the Burgermeister and Uflacker references, they did not appear to teach this feature, but noted that Applicant's specification may not provide sufficient support for an amendment that would claim this feature. However, although Examiner suggested filing a continuation-in-part more fully describing all features of

Application No.: 10/522,889

Statement of Substance of the Interview held on May 15, 2009

the invention to increase the likelihood of finding allowable subject matter, upon review Applicant identified direct support in the specification for amendment that Applicant considers would claim the above mentioned feature. Applicant therefore will submit a detailed next Amendment to Examiner to justify claiming this feature on the basis of the current specification.

Respectfully submitted,

FREEFLOW MEDICAL LIMITED

By



Rudy Hengelmolen

Mobile: +44 (0)7801 703889

E-mail: [rudym.hengelmolen@gmail.com](mailto:rudym.hengelmolen@gmail.com)